BALANCE CALIBRATION LOG

BENCH 2A

Margin of Error for FX-2000 is +/- 0.02g

Margin of Error for Explorer Pro is +/- 0.0002g

Balance A = FX-2000

Balance B = Explorer Pro (Model # EP214C)

Weekly Weighted Calibration

BALANCE A				BALANCE B		
Date	Chemist Initials	Standard Weight (g)	Result Weight (g)	Date	Chemist Initials	Standard Weight (g)
		1				1
		2				2
		5				5
		10				10
		20				20
		50				50
		1				1
		2				2
		5				5
		10				10
		20				20
		50				50
		1				1
		2				2
		5				5
		10				10
		20				20
		50				50
		1				1
		2				2
		5				5
		10				10
		20				20
		50				50

Result Weight (g)

BALANCE CALIBRATION LOG		Margin of Error for FX-2000 is +/- 0.02g
BENCH 2A		Margin of Error for Explorer Pro is +/- 0.0002g

Balance A = FX-2000
Balance B = Explorer Pro (Model # EP214C)

QC Procedure
1.) Internally Calibrate Daily (as used)
2.) Calibrate against weights weekly
3.) Routinely calibrate by outside contractor - record date in log

Daily Internal Calibration			Daily Internal Calibration	
Dates	Chemist Initials		Dates	Chemist Initials